



Measuring Instruments For Wireless Communications

Inspected by	Approval by
윤병민	김광국

Inspection Record

Product		Inspection	
Description	TEM Cell	Date of Test	2023-06-05
Model No.	TC-5062C	Temperature	25 ±2℃
Serial No.	5062C000832	Humidity	< 55%

Equipment Used :

Model Number	Model Description	Serial No.
MS2663C	Spectrum Analyzer	6200050489
TC-94017A	Mini SG KIT	-
TC-92020A	Dipole Antenna(900MHz~1.8GHz)	-
TC-92040A	Dipole Antenna(2.4GHz)	-
TC-92030A	Sleeve Antenna(5.8GHz)	-


Measurement Report

NO.	S/N	Frequency	Results			Remarks
			Min	Actual	Max	
1	5062C000832	900MHz	-80.0dB	-90.8dB	-	Pass
		1800MHz	-80.0dB	-89.0dB	-	Pass
		2400MHz	-70.0dB	-85.2dB	-	Pass
		5800MHz	-60.0dB	-80.0dB	-	Pass

Test Report

- V.S.W.R. Graph

SERIAL NUMBER : 5062C000832



- V.S.W.R. TEST SETUP

< 1.7 , 100MHz ~ 6GHz

