



Measuring Instruments For Wireless Communications

Inspected by	Approval by
박찬웅	김광주

Inspection Record

Product		Inspection	
Description	Pneumatic Shield Box	Date of Test	2023-12-04
Model No.	TC-5915APU	Temperature	25 ±2℃
Serial No.	5915APU000102	Humidity	< 55%

1. Shield Effect

Equipment Used:

Model Number	Model Description	Serial No.
-	Vivaldi Antenna (0.9GHz to 12GHz)	-
E5071C	Network Analyzer (30kHz ~ 20GHz)	MY46632264

Measurement Report

S/N	Frequency	Results			Remarks
		Min	Actual		
5915APU000102	900MHz	-60dB	-101dB	Pass	PASS
	1800MHz	-60dB	-104dB	Pass	
	2400MHz	-70dB	-106dB	Pass	
	5800MHz	-60dB	-104dB	Pass	
	7GHz	-60dB	-98dB	Pass	
	8GHz	-60dB	-98dB	Pass	
	9GHz	-60dB	-86dB	Pass	
	10GHz	-60dB	-85dB	Pass	
	11GHz	-60dB	-79dB	Pass	
	12GHz	-60dB	-77dB	Pass	

* The shield effectiveness above is measured using TESCOM test module